ProEdge
DENTAL WATER LABS

ProEdgeDental.com | 888.843.3343 | 7042 S Revere Pkwy #400, Centennial, CO 80112

R2A DENTAL WATER SAFETY REPORT

ProEdge Dental Water Lab Technicians approve the following results accurate and complete utilizing gold-standard testing and laboratory methods.

Colton Knepp
Quality Assurance

Hillary Sinclair
Customer Delight

Natalie Lenser DDS

Modesto, CA	Phone number	Test date	Report date
	209-571-7283	01/09/2025	01/15/2025

Waterline Treatment Info

Source water	Shocks used	Shock date	Treatment used
Bottled Distilled	Liquid Ultra (Pink Stuff)	01/02/2025	ICX

SAMPLES SUBMITTED / LABORATORY TEST RESULTS (CFU/mL)

ProEdge sample #	Operatory #	Sample source	CDC compliance	Microbial count	0	5-200	201-500	501-2,500	2,501-10,000	10,001-40,000	40,001-100,000	TNTC
490	Blue	AW Syringe	Pass	0	●							
491	Yellow	Highspeed handpiece	Pass	0	●							
492	Green	AW Syringe	Pass	0	●							
493	Brown	Highspeed handpiece	Pass	0	●							
494	Green Op	AW Syringe	Pass	0	●							
495	Yellow Op	AW Syringe	Pass	0	●							
496	Blue Op	Highspeed handpiece	Pass	0	●							